Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0237824	(X3) Date Survey Completed 11/01/2022
Name of Provider or Supplier White Oak Family Physicians, Pa	**Street Address, City, State** 550 White Oak Street, Asheboro, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D1001	CERTIFICATE OF WAIVER TESTS CFR(s): 493.15(e) Laboratories eligible for a certificate of waiver must-- (1) Follow manufacturers' instructions for performing the test; and (2) Meet the requirements in subpart B, Certificate of Waiver, of this part. This STANDARD is not met as evidenced by: Based on review of manufacturer's instructions, observation, and review of the October 2022 rapid strep patient log 11/1/22, the laboratory failed to follow manufacturer's instructions for performance of the Henry Schein One Step + Strep A Dipstick test. The laboratory failed to discard an expired rapid strep kit, and the kit was used to test three patients. Findings: Review of manufacturer's instructions for the Henry Schein One Step + Strep A Dipstick test revealed "... PRECAUTIONS ... Do not use after expiration date. ... STORAGE AND STABILITY ... Do not use beyond the expiration date. ..." During a tour of the facility at approximately 3:10 p.m., the surveyor observed an open Henry Schein One Step + Strep A Dipstick kit lot #STA0102019 with expiration date 9/30/22 in a cabinet in the nurses' station, available for use. Review of the October 2022 patient log revealed the expired strep kit was used to perform testing on three patients: 1. #1079190 was tested 10/3/22; 2. #857901 was tested 10/7/22; 3. #1155920 was tested 10/28/22.