

|  |  |   |
|--|--|---|
| <b>Statement of Deficiencies</b>   | <b>(X1) Provider/Supplier/CLIA Identification Number</b><br><br>34D0237824   | <b>(X3) Date Survey Completed</b><br><br>09/26/2024 |
| <b>Name of Provider or Supplier</b><br><br>White Oak Family Physicians, Pa   | <b>Street Address, City, State</b><br><br>550 White Oak Street, Asheboro, NC |   |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. |  |   |

| <b>(X4) ID Prefix Tag</b> | <b>Summary Statement of Deficiencies</b>   |
|---------------------------|--|
| <b>D5417</b>              | <p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT<br/>CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by:<br/>Based on observation and interview with TP (testing personnel) #1 on 9/26/24, the laboratory failed to discard KOH (potassium hydroxide) that exceeded the expiration date. Findings: During a tour of the laboratory at approximately 3:10 p.m., the surveyor observed one bottle of KOH (lot #3234-00 with expiration date 3/31/24) in the Side A microscope area, available for use. During interview at approximately 3:10 p.m., TP #1 confirmed the KOH was expired and should have been discarded.</p> |