

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0238143	(X3) Date Survey Completed 05/30/2019
Name of Provider or Supplier Central Carolina Dermatology Clinic, Inc	Street Address, City, State 4010 Mendenhall Oaks Parkway, High Point, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of 2017, 2018, and 2019 MLE (Medical Laboratory Evaluation) proficiency testing records and staff interview on 5/30/19, the laboratory failed to evaluate all ungraded proficiency testing results received to ensure corrective action was taken and documented as required. Findings: 1. Review of 2017 MLE 3rd event revealed 1 of 2 DTM (Dermatophyte Test Medium) results were ungraded with no evaluation documented. 2. Review of 2018 MLE 3rd event revealed 4 of 5 DTM results were ungraded with no evaluation documented. 3. Review of 2019 MLE 1st event revealed 2 of 5 DTM results were ungraded with no evaluation documented. During interview at approximately 1:00 p.m., histotech #1 stated that she printed the data summary pages for each proficiency testing event, but the laboratory director failed to evaluate the results to determine whether corrective action was needed.</p>