

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  34D0238795	<b>(X3) Date Survey Completed</b>  03/13/2020
<b>Name of Provider or Supplier</b>  Caswell Family Medical Center	<b>Street Address, City, State</b>  439 Us Highway 158 West, Yanceyville, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6098</b>	<p><b>LABORATORY DIRECTOR RESPONSIBILITIES</b> CFR(s): 493.1445(e)(8)</p> <p>The laboratory director must ensure that reports of test results include pertinent information required for interpretation.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory test reports and interview with laboratory director (LD) 3/13/20, the LD failed to ensure patient test reports for urine microscopic and wet prep testing included pertinent information required for interpretation. Findings: Review of patient test report, date of birth (DOB) 9/10/1979, for the wet prep test revealed the reference range for Clue Cells was recorded as saline-saline. Review of patient test report, DOB 8/15/1958, for the urine microscopic test revealed no reference ranges for the following: White Blood Cell (WBC), Red Blood Cell (RBC), Epithelial Cells, Bacteria, Yeasts, Crystals or Casts. During exit interview at approximately 11:45 a.m. the LD confirmed the test reports failed to include pertinent information required for interpretation. She stated they did not realize the Clue Cells reference range showed up as saline-saline. She also stated the urine microscopic reference ranges were on previous reports, but they must have dropped off when a computer update took place.</p>