Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0720068	(X3) Date Survey Completed 11/02/2023
Name of Provider or Supplier Novant Health Mintview Ob/Gyn Bsllantyne	**Street Address, City, State** 10905 Providence Road West, Suite 260, Charlotte, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6046	TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8) (b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently. This STANDARD is not met as evidenced by: Based upon review of competency documentation, review of 2022 and 2023 API proficiency testing documentation and interview with the TC (Technical Consultant) on 11/2/23, the TC failed to assess the competency of 1 of 9 TP (Testing Personnel) performing non waived testing in the laboratory. Findings: Review of competency documentation revealed the absence of any competency assessment of TP #8 for use of the BD Affirm analyzer. Competency assessments completed for TP #8 included only waived test systems. Review of 2022 and 2023 API proficiency testing documentation revealed that TP #8 participated in the Second Microbiology Event of 2023. In interview at approximately 1:45 p.m., the TC stated she was unaware that TP #8 was performing non waived testing in the laboratory. The TC consulted the Office Manager and confirmed that TP #8 has been performing patient testing utilizing the BD Affirm analyzer.