

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0864770	(X3) Date Survey Completed 07/20/2020
Name of Provider or Supplier Harnett County Health Department	Street Address, City, State 307 W Cornelius Harnett Boulevard, Lillington, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on desk review of CMS (Centers for Medicare and Medicaid Services) Casper report number 153D and 155D and the 2020 AAB (American Association of Bioanalysts) proficiency testing results 7/20/20, the laboratory failed to successfully participate in proficiency testing for Bacteriology. See the deficiency cited at D2028.
D2028	BACTERIOLOGY CFR(s): 493.823(e) Failure to achieve an overall testing event score of satisfactory performance for two

	consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on desk review of CMS(Centers for Medicare and Medicaid Services) Casper Reports 153D and 155D and desk review of 2020 AAB (American Association of Bioanalysts) proficiency testing results 7/20/20, the laboratory failed to achieve satisfactory performance for Bacteriology in two consecutive testing events, resulting in unsuccessful performance. Findings: 1. Desk review of CMS Casper report 155D and 2020 AAB proficiency testing results revealed the laboratory provided unacceptable responses for 2 of 5 Genital culture samples, resulting in a score of 60% for Bacteriology on the 2020 Nonchemistry 1st event. 2. Desk review of CMS Casper report 155D and 2020 AAB proficiency testing results revealed the laboratory provided unacceptable responses for 2 of 5 Gram stain samples, resulting in a score of 60% for Bacteriology on the 2020 Nonchemistry 2nd event.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on desk review of CMS(Centers for Medicare and Medicaid Services) Casper reports 153D and 155D and desk review of 2020 AAB(American Association of Bioanalysts) proficiency testing results 7/20/20, the laboratory director failed to provide overall management and direction to ensure successful proficiency testing participation. See the deficiency cited at D6016.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on the desk review of CMS(Centers for Medicare and Medicaid Services) Casper reports 153D and 155D and desk review of 2020 AAB(American Association of Bioanalysts) proficiency testing results 7/20/20, the laboratory director failed to ensure successful participation in proficiency testing as required in Subpart H. Findings: 1. Desk review of CMS Casper report 155D and 2020 AAB proficiency testing results revealed the laboratory provided unacceptable responses for 2 of 5 Genital culture samples, resulting in a score of 60% for Bacteriology on the 2020 Nonchemistry 1st event. 2. Desk review of CMS Casper report 155D and 2020 AAB proficiency testing results revealed the laboratory provided unacceptable responses for

2 of 5 Gram stain samples, resulting in a score of 60% for Bacteriology on the 2020 Nonchemistry 2nd event.