

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0918362	(X3) Date Survey Completed 10/29/2024
Name of Provider or Supplier A Woman's View	Street Address, City, State 915 Tate Boulevard Se, Suite 170, Hickory, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory director on 10/29/2024, the laboratory failed to discard materials that had exceeded their expiration dates and were available for use in patient testing. Findings: 1. At approximately 3:00 P.M., the surveyor observed one (1) Vacuette Red Top blood collection tube with Lot Number B2303338 and expiration date 2024-08-30 in a tray in the Phlebotomy Collection Station inside the laboratory. 2. At approximately 3:00 P.M., the surveyor observed four (4) QFT LabCorp blood collection kits with Lot Number 2309280177 and expiration date 09/30/24 in a tray in the Phlebotomy Collection Station inside the laboratory. The expired blood collection tubes were given to laboratory director who, during an interview at 3:00 P.M., confirmed that they had exceeded their expiration dates. The expired tubes were immediately discarded in the laboratory's biohazard container.</p>