

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0950270	(X3) Date Survey Completed 11/10/2021
Name of Provider or Supplier Cary Skin Center, Pa	Street Address, City, State 200 Wellesley Trade Lane, Cary, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory verification of accuracy records and interview with testing personnel (TP) #4 11/10/21, the laboratory failed to verify the accuracy of the Mart I Immunostain at least twice annually in 2019, 2020, and 2021, a period of approximately 3 years in which verification of accuracy was not performed. Findings: Review of laboratory verification of accuracy records revealed no documentation the laboratory had performed a twice annual verification of accuracy for the Mart I Immunostain in 2019, 2020, and 2021. During interview at approximately 2:30 p.m. TP #4 confirmed the laboratory did not perform a twice annual verification of accuracy for the Mart I Immunostain in 2019, 2020 and 2021. He stated when he took over the position in March of 2021 a twice annual verification of accuracy was not being performed for the Mart I Immunostain, so he did not know that it was needed.</p>