

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0975613	(X3) Date Survey Completed 10/26/2022
Name of Provider or Supplier Coastal Dermatology And Surgery Center, Pa	Street Address, City, State 2504 Delaney Avenue, Wilmington, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2003	<p>ENROLLMENT CFR(s): 493.801(a)(2)(ii)</p> <p>For those tests performed by the laboratory that are not included in subpart I of this part, a laboratory must establish and maintain the accuracy of its testing procedures, in accordance with 493.1236(c)(1)</p> <p>This STANDARD is not met as evidenced by: Based on review of 2020, 2021 and 2022 laboratory records and interview with laboratory director (LD) 10/26/22, the laboratory failed to verify the accuracy of the Mohs testing performed at least twice annually in 2020 and 2021 and at least once in 2022, a period of approximately 30 months in which a verification of accuracy was not performed. Findings: Review of 2020, 2021 and 2022 laboratory records revealed no documentation of a verification of accuracy for the Mohs testing. Interview with LD at approximately 12:00 p.m. confirmed the laboratory did not perform a verification of accuracy for the Mohs testing for approximately 30 months. The LD stated there was a large turn over in staff and sending slides for review was overlooked.</p>