Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0992791	(X3) Date Survey Completed 04/03/2025
Name of Provider or Supplier Americare Health Plaza Medical Center	**Street Address, City, State** 1805 Milton Road, Charlotte, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) (d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on observation and interview with TP (testing personnel) #1 on 4/3/25, the laboratory failed to discard supplies that were past their expiration dates for use. During a tour of the laboratory at approximately 2:00 pm, the surveyor observed the following expired materials in a drawer, available for use: 1. 5 stool collection kits with 3 Para-Pak vials in each kit, including one C&S (culture & sensitivity) vial, one SAF (sodium acetate-acetic acid-formalin) vial, and one Clean vial. Each kit contained one or more expired vials which included two C&S vials lot 437730-P expiration 12/23/23, three SAF vials lot 417540-P expiration 8/17/23, and three SAF vials lot 413340-P expiration 1/17/25. 2. 2 Total-Fix MCC (Medical Chemical Corporation) collection vials for Ova and Parasite detection, lot 3645-00 expiration 07-24. 3. 3 orange Aptima Multi-test Swabs with Transport Media (STM), lot 331072 expiration 2/29/24. During an interview at approximately 2:15 pm, TP #1 stated the the supplies are not used by the laboratory.