Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D0994130	(X3) Date Survey Completed 07/26/2019
Name of Provider or Supplier Piedmont Plastic Surgery & Dermatology - Gastonia	**Street Address, City, State** 959 Cox Road, Gastonia, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on review of laboratory policies, review of 2017, 2018 and 2019 proficiency testing (PT) records and laboratory manager interview 7/26/19, the laboratory failed to verify the accuracy of Mohs testing at least twice a year. Review of laboratory policy, "Biannual Proficiency Testing" revealed "Scope/Purpose..The quality assurance program for verifying the accuracy and reliability of the testing results in the Mohs laboratory. Four cases will be randomly selected twice a year for review." Review of PT records for 2017, 2018 and 2019 revealed the laboratory failed to perform biannual proficiency testing to verify the accuracy of Mohs testing in 2018. Interview with laboratory manager at approximately 12:00 p.m. confirmed the laboratory failed to verify the accuracy of Moths testing at least twice a year.