

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D1024517	(X3) Date Survey Completed 06/28/2019
Name of Provider or Supplier Piedmont Plastic Surgery & Dermatology - Hickory	Street Address, City, State 315 19th Street Se, Hickory, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation, records review, and staff interview on 6/28/19, the laboratory failed to discard materials that had exceeded its expiration date and was available for use. Findings: 1. At 9:20 a.m. the surveyor observed a bottle HealthLink 20% KOH, Lot #: 171211, that had expired on 2018-05-01 located on the north hallway on the shelf above the microscope. 2. The Mohs Tech indicated during interview at 9:20 a. m., that the Nursing Clinical Team Manager is responsible for this testing area and the manager failed to notice the expired material. 3. The review of the KOH Testing log revealed that approximately 28 patient test were performed using the expired HealthLink 20% KOH.</p>