

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D1092661	(X3) Date Survey Completed 02/22/2024
Name of Provider or Supplier Heag Pain Management Center	Street Address, City, State 1305 West Wendover Avenue, Suite A, Greensboro, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5467	<p>CONTROL PROCEDURES CFR(s): 493.1256(d)(9)(g)</p> <p>Unless CMS Approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing, the laboratory must-- When using calibration material as a control material, use calibration material from a different lot number than that used to establish a cut-off value or to calibrate the test system. (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based upon interview with TP#1 (Testing Personnel) on 2/22/24, the laboratory failed to prepare LC/MS (liquid chromatography/mass spectrometry) quality control and calibration materials from different standard stock solutions. Findings: During interview at approximately 1:35 p.m., TP#1 stated that both quality control and calibration materials are prepared from the same standard stock solution.</p>