

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D2003984	(X3) Date Survey Completed 05/03/2019
Name of Provider or Supplier Masonboro Urgent Care, Pc	Street Address, City, State 6132 Carolina Beach Rd, Suite 8, Wilmington, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5775	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities.</p> <p>This STANDARD is not met as evidenced by: Based on the absence of documentation and interview with the practice / laboratory manager on 5/3/19 at 9:40 a.m., the laboratory failed to perform and maintain documentation of Hemoglobin (Hgb) and Hematocrit (Hct) comparison testing between the Medonic M Series and Abbott i-STAT hematology analyzers at least twice a year. The laboratory manager and director acknowledged during interview at 9:45 a.m. that they failed to perform the required Hgb and Hct twice a year method comparison between the two analyzers. They further acknowledged that the failure to perform the required comparison testing was an oversight and not an intentional act.</p>