

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  34D2024372	<b>(X3) Date Survey Completed</b>  05/15/2026
<b>Name of Provider or Supplier</b>  Rainbow Pediatrics - Hoke County Location	<b>Street Address, City, State</b>  142 Paraclete Drive, Suite 102, Raeford, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5215</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on review of 2025 American Proficiency Institute (API) proficiency testing (PT) records, review of laboratory procedure, lack of documentation and interview with testing personnel (TP) #1, 05/15/25, the laboratory failed to verify the accuracy of PT results when zero scores were obtained for nonparticipation. Findings: Review of API 2025 Hematology/Coagulation - 3rd Event records revealed the laboratory scored zero for all samples and all analytes due to nonparticipation. Review of laboratory procedure revealed "If the PT results is returned as "not graded", review it against the summary report from the PT agency to ensure that the answers submitted would have been acceptable....". The procedure fails to include a review if the laboratory fails to participate in the PT event. Review of laboratory PT records revealed the laboratory tested the 2025 3rd Event PT samples. The records failed to include a review of the results obtained to ensure the accuracy of the PT testing. The records also failed to include a review of the zero scores by the laboratory director /technical consultant. Interview with TP #1 at approximately 12:00 p.m. confirmed the laboratory director/technical consultant did not review the 2025 3rd Event PT scores. They also confirmed no review of the PT summary was performed to ensure the accuracy of the PT testing.</p>