

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  34D2068467	<b>(X3) Date Survey Completed</b>  07/23/2024
<b>Name of Provider or Supplier</b>  Laboratory Corporation Of America Holdings	<b>Street Address, City, State</b>  1710 Kernersville Medical Parkway Suite 116, Kernersville, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with technical supervisor (TS) #2 on 7/23/2024, the laboratory failed to discard materials that had exceeded their expiration dates and were available for use in patient testing. Findings: 1. At approximately 12:30 P.M., the surveyor observed 14 Becton Dickinson (BD) Blue top Vacutainer blood collection tubes with Lot Number 3257668 and expiration date 2024-06-30 in a cup in the Phlebotomy Collection Station inside the laboratory. 2. At approximately 12:30 P. M., the surveyor observed 1 BD Green top Vacutainer blood collection tube with Lot Number 2347082 and expiration date 2024-04-30 in a cup in the Phlebotomy Storage closet in the laboratory. The expired BD blood collection tubes were given to TS #2 who, during an interview at 12:30 P.M., confirmed that they had exceeded their expiration dates. The expired tubes were immediately discarded in the laboratory's biohazard container.</p>