

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D2080229	(X3) Date Survey Completed 08/16/2019
Name of Provider or Supplier Dermatology And Skin Surgery Center At, The	Street Address, City, State 7200 Creedmoor Road, Suite 104, Raleigh, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory policy, review of 2017, 2018 and 2019 laboratory proficiency records and interview with laboratory technician 8/16/19, the laboratory failed to verify the accuracy of the Mohs testing at least twice annually. Findings: 1. Review of laboratory procedure "Proficiency Testing...Mohs Micrographic Surgery Skin Specimens....Periodically, the tech or Risk Manager will send a collection of every 30th case containing the original slides, out for a microscopic examination by a Board Certified Pathologist,...Results of each Proficiency Test will be entered in a log and kept in the laboratory management manual,..." 2. Review of 2017, 2018 and 2019 laboratory proficiency records revealed the laboratory failed to verify the accuracy of the Mohs testing at least twice annually from approximately 9/17 until 10/18, a period of approximately 12 months. Interview with laboratory technician at approximately 12:00 p.m. confirmed there was no documentation to verify the accuracy of the Mohs testing from 9/17 until 10/18.</p>