

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D2168553	(X3) Date Survey Completed 07/15/2021
Name of Provider or Supplier Wake Skin Cancer Center	Street Address, City, State 11640 North Park Dr - Suite 200, Wake Forest, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory verification of accuracy records and interview with histotechnician 7/15/21, the laboratory failed to verify the accuracy of the Mohs testing and the Melanoma antigen recognized by T cells (MART-1) testing at least twice annually. 1. The laboratory failed to verify the accuracy of the Mohs testing twice annually in 2020. Findings: Review of laboratory verification of accuracy records revealed the laboratory had verified the accuracy of the Mohs testing in February of 2020, but failed to perform a verification of accuracy twice annually as required. Interview with histotechnician at approximately 10:30 a.m. confirmed the laboratory had failed to perform a verification of accuracy for the Mohs testing at least twice in 2020. She stated they thought they would only have to do a verification of accuracy once a year. 2. The laboratory failed to verify the accuracy of the MART-1 testing twice annually in 2020. Findings: Review of laboratory verification of accuracy records revealed no documentation the laboratory had verified the accuracy of the MART-1 testing twice annually in 2020. Interview with histotechnician at approximately 10:30 a.m. confirmed the laboratory had failed to perform a verification of accuracy for the MART-1 testing twice annually in 2020. She stated they did not realize they would have to perform a verification of accuracy for the MART-1 testing.</p>