

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 34D2184524	(X3) Date Survey Completed 06/07/2023
Name of Provider or Supplier Atlantic Dermatology, Pa	Street Address, City, State 720 Jefferson Sreet, Whiteville, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of 2020, 2021, and 2022 laboratory records and interview with the histology technician 6/7/23, the laboratory failed to verify the accuracy of the Mohs histopathology testing at least twice a year during 2020, 2021, and 2022. Review of 2020, 2021, and 2022 laboratory records revealed the laboratory had participated in peer reviews of Mohs surgery cases to verify the accuracy of their Mohs histopathology testing. The peer reviews were performed only once per year in 2020, 2021, and 2022. During interview at approximately 11:00 a.m., the histology technician stated that the laboratory director had also participated in peer reviews at the laboratory's sister facility, so they were unaware he needed to participate at least twice a year at this location.</p>