

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  34D2200341	<b>(X3) Date Survey Completed</b>  11/18/2024
<b>Name of Provider or Supplier</b>  Atrium Health Wfb General Dermatology Laboratory -	<b>Street Address, City, State</b>  5826 Samet Drive Suite 103, High Point, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with the histotechnician 11/18/24, the laboratory failed to discard supplies that had exceeded their expiration dates. Findings: During a tour of the laboratory at approximately 2:00 p.m., the surveyor observed the following expired supplies in the microscope room adjacent to the laboratory, available for use: 1. 1 bottle of Delasco 1% Toluidine Blue Aqueous Solution, lot #K20CG2 with expiration date 12/31/22; 2. 1 set of Fisherbrand Hema 3, including: a. 1 bottle of Hema 3 Fixative, lot #100035 with expiration date 11/2021; b. 1 bottle of Hema 3 Solution I, lot #100036 with expiration date 2/2022; c. 1 bottle of Hema 3 Solution II, lot #100037 with expiration date 2/2022. During interview at approximately 2:15 p. m., the histotechnician stated she was unaware the supplies were in that location and she was unsure what they were used for.</p>