

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 35D0408585	(X3) Date Survey Completed 02/03/2026
Name of Provider or Supplier Langdon Prairie Health	Street Address, City, State 909 2nd St, Langdon, ND	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5775	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and staff interview, the laboratory failed to twice annually compare and evaluate test results using different methodologies for non-waived chemistry analytes (sodium, potassium, chloride, total CO2 [carbon dioxide], glucose, BUN [blood urea nitrogen], and creatinine) for 1 of 1 year reviewed (2025). Findings include: 1. Observation of the laboratory, beginning at 7:49 a.m. on 02/03/26, indicated the laboratory performed sodium, potassium, chloride, total CO2, glucose, BUN, and creatinine on the i-STAT and Siemens Dimension EXL 200 (Dimension). 2. Reviewed the afternoon of 02/03/26, the 2025 test comparison records lacked evidence of i-STAT and Dimension method comparisons. Upon request, the laboratory failed to provide evidence of comparisons in 2025 for sodium, potassium, chloride, total CO2, glucose, BUN, and creatinine with the i-STAT and Dimension. 3. During interview at approximately 4:30 p.m. on 02/03/26, a supervisory staff member (#1) confirmed the laboratory performed sodium, potassium, chloride, total CO2, glucose, BUN, and creatinine patient testing and did not perform comparisons on the i-STAT and Dimension.</p>