

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 36D0667729	(X3) Date Survey Completed 03/08/2022
Name of Provider or Supplier Pediatric Place Inc	Street Address, City, State 3690 Orange Place #100, Beachwood, OH	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2020	<p>BACTERIOLOGY CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review and an interview with the Laboratory Director, the laboratory failed to to attain an overall Bacteriology proficiency testing (PT) event score of at least 80%. This deficient practice had the potential to affect 25 out of 25 patient throat cultures performed between 02/26/2021 and 06/07/2021. Findings Include: 1. Review of the laboratory's 2021 Wisconsin State Laboratory Hygiene (WSLH) PT records revealed an overall Bacteriology (throat culture) testing event score of 60% for the first PT event of 2021. 2. The Laboratory Director confirmed that the laboratory did not attain an overall WSLH throat culture PT event score of at least 80% for the first PT testing event in 2021. The interview occurred on 03/08/2022 at 10:10 AM.</p>