

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 36D1002988	(X3) Date Survey Completed 02/20/2019
Name of Provider or Supplier Ibrahim Zayneh Md Llc	Street Address, City, State 2127 25th Street, Portsmouth, OH	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's proficiency testing (PT) documentation, and an interview with the Lab Director (LD) the laboratory failed to verify the accuracy of all unregulated analytes test procedures performed, at least twice annually. All patients have the potential to be affected. Findings Include: 1. Review of 2 of 2 of the laboratory's 2017 PT documents, and 2 of 2 of the laboratory's 2018 PT documents provided on the date of inspection did not find a reviewers signature and date, number of stages for each case, dates sent and received back, or a LD reviewed signature and date. 2. The LD confirmed the 2017 and 2018 PT documents were missing the above information and there was no way to verify slides had been sent for peer review. The interview occurred on 02/20/2019 at 10:13 AM.</p>