

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 36D1036910	(X3) Date Survey Completed 08/06/2026
Name of Provider or Supplier Dillon Dermatology	Street Address, City, State 1037 Conneaut, Suite 201, Bowling Green, OH	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2000	ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801 Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994. This CONDITION is not met as evidenced by: Based on record review and an interview with Mohs Technician #1 , the laboratory failed to enroll in a proficiency testing (PT) program for the moderate complexity potassium hydroxide (KOH) microscopic skin test in the subspecialties of Mycology and Parasitology. This deficient practice had the potential to affect two out of two subspecialties tested from 10/01/2024 through 08/06/2026. Findings Include: 1. Surveyor review of the KOH policies and procedures failed to find mention of proficiency testing enrollment for the KOH test in the subspecialties of Mycology and Parasitology 2. An interview with Mohs Technician #1, on 08/06/2026 at 12:44 PM, confirmed that the laboratory failed to enroll in a proficiency testing program for the moderate complexity KOH test in the subspecialties of Mycology and Parasitology. 3. This deficient practice had the potential to affect 90 out of 90 patient test results from 10/01/2024 through 08/06/2026.