

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 36D2111241 | (X3) Date Survey Completed 02/21/2018 |
| Name of Provider or Supplier Universal Dermatology | Street Address, City, State 7100 Graphics Way, Suite 3000, Lewis Center, OH | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D5217 | <p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on policy and procedure review, and an interview with the Technical Supervisor (TS), the laboratory failed to blindly verify the accuracy of the Mohs procedure, not listed in Subpart I, at least twice annually. Findings Include: 1. Review of the laboratory's policy and procedure found a document titled "PROFICIENCY TESTING MOHS MICROGRAPHIC SURGERY SKIN SPECIMENS" containing the following statement: "Semi-annually, the tech or Risk Manager will send two cases containing the original slides, label it with only surgical case number, and send it our for a microscopic examination by a Board Certified Pathologist." 2. An interview with the TS confirmed no Mohs slides were sent for verification of accuracy in 2016. The interview occurred on 02/21/2018.</p> |