

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  36D2197506	<b>(X3) Date Survey Completed</b>  08/10/2021
<b>Name of Provider or Supplier</b>  Drugstore At O'Bleness, The	<b>Street Address, City, State</b>  934 E State St, Athens, OH	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D1002</b>	<p><b>REPORTING OF SARS-CoV-2 TEST RESULTS</b></p> <p>During the Public Health Emergency, as defined in 400.200 of this chapter, each laboratory that performs a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 (hereinafter referred to as a "SARS-CoV-2 test") must report SARS-CoV-2 test results to the Secretary in such form and manner, and at such timing and frequency, as the Secretary may prescribe.</p> <p>This CONDITION is not met as evidenced by: Based on record review and an interview with the Laboratory Director found that the laboratory failed to report all SARS-CoV-2 test results as required for five out of six patients tested and reviewed from January 27, 2021 through May 10, 2021. Findings Include: 1. Review of the laboratory's "Facility Testing Procedure" policy and procedure, provided for inspection of the SARS-CoV-2 testing and reporting documentation found the following statement: "...11....a. Confirmed test results must be reported to the local health department within 24 hours." 2. The laboratory's SARS-CoV-2 patient testing and result reporting documentation was reviewed from 01/27/2021 through 05/10/2021 and found 1 positive patient COVID-19 tested and reported to the local and Ohio Department of Health departments. Further review revealed five out of five negative patient COVID-19 test results were not reported. 3. The Inspector requested the laboratory's evidence that the above mentioned negative COVID-19 results were reported to the Ohio Department of Health from the Laboratory Director. The Laboratory Director confirmed on 08/10/2021 at 9:44 AM, that the laboratory did not report five out of five patient COVID-19 negative test results identified during the remote inspection and were unable to provide the requested documentation.</p>