Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D0676326	(X3) Date Survey Completed 06/20/2022
Name of Provider or Supplier Southwestern Dermatology	**Street Address, City, State** 8315 S Walker Ave, Oklahoma City, OK	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 06/20/2022. The laboratory was found in compliance with a standard-level deficiency cited. The findings were reviewed with the laboratory manager and office manager at the conclusion of the survey.
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on a review of records and interview with the laboratory manager, the laboratory failed to verify the accuracy of slide interpretations at least twice annually. Findings include: (1) On 06/20/2022 at 01:45 pm, the laboratory manager stated the laboratory performed microscopic interpretations of H&E (Hematoxylin and Eosin) stained slides from frozen section skin biopsies and tissues removed during Mohs surgery. The tissue would then be observed microscopically; (2) A review of records for testing performed from December 2020 through the day of the survey revealed the accuracy of the slide interpretations for frozen section biopsies and Mohs surgical specimens had not been verified twice annually. The interpretations had not been verified for accuracy between: (a) 12/30/2020 and 08/10/2021 (b) 08/10/2021 and 01/10/2022 (3) The records were reviewed with the laboratory manager who stated on 06/20/2022 at 02:52 pm the slide interpretations had not been verified for accuracy twice during 2021.