

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 37D0699027	<b>(X3) Date Survey Completed</b> 12/27/2021
<b>Name of Provider or Supplier</b> Skin Clinic, The	<b>Street Address, City, State</b> 2410 North Commerce, Ardmore, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The recertification survey was performed on 12/27/2021. The findings were reviewed with the histotechnician at the conclusion of the survey. The laboratory was found in compliance with standard-level deficiencies cited.
<b>D5473</b>	<p><b>CONTROL PROCEDURES</b> CFR(s): 493.1256(e)(2)(g)</p> <p>(e) For reagent, media, and supply checks, the laboratory must do the following: (e) (2) Each day of use (unless otherwise specified in this subpart), test staining materials for intended reactivity to ensure predictable staining characteristics. Control materials for both positive and negative reactivity must be included, as appropriate. (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on a review of records and interview with the histotechnician, the laboratory failed to document the reactivity of the H&amp;E (Hematoxylin &amp; Eosin) stain each day of testing for four of 52 days of patient testing. Findings include: (1) On 12/27/2021 at 11:15 am, the histotechnician stated the laboratory performed microscopic interpretations of histology (Mohs and skin biopsy) specimens that had been stained with H&amp;E (Hematoxylin &amp; Eosin) stain; (2) The surveyor reviewed test records for 52 days of patient testing (microscopic interpretations) performed from 01/06/2021 through 12/16/2021. There was no evidence that the reactivity of the stain had been observed for acceptability for four (03/12/2021, 05/28/2021, 10/22/2021, and 11/05/2021) of the 52 days; (3) The surveyor reviewed the findings with the histotechnician who stated on 12/27/2021 at 12:23 pm, the reactivity of the stain was observed for acceptability on each patient slide, but had not been documented as indicated above.</p>