

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D0699027	(X3) Date Survey Completed 10/11/2023
Name of Provider or Supplier Skin Clinic, The	Street Address, City, State 2410 North Commerce, Ardmore, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 10/11/2023. The laboratory was found in compliance with standard-level deficiencies cited. The findings were reviewed with the histology technician during an exit conference performed at the conclusion of the survey.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with the histology technician, the laboratory failed to ensure that one of one expired reagent was not available for use. Findings include: (1) Observation of the laboratory on 10/11/2023 at 10:19 am identified the following reagent stored in the flammable cabinet that appeared to be available for use: (a) One bottle of Mercedes Scientific Platinum Line Xylenes, lot #2026901 with an expiration date of 03/31/2023. (2) The finding was reviewed with the histology technician who stated on 10/11/2023 at 10:23 am, the reagent was available for use.</p>