

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D0699027	(X3) Date Survey Completed 07/30/2025
Name of Provider or Supplier Skin Clinic, The	Street Address, City, State 2410 North Commerce, Ardmore, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 07/30/2025. The laboratory was found in compliance with a standard-level deficiency cited. The findings were reviewed with the histotechnology supervisor and medical assistance at the conclusion of the survey.
D5473	<p>CONTROL PROCEDURES CFR(s): 493.1256(e)(2)(g)</p> <p>(e)(2) Each day of use (unless otherwise specified in this subpart), test staining materials for intended reactivity to ensure predictable staining characteristics. Control materials for both positive and negative reactivity must be included, as appropriate.</p> <p>This STANDARD is not met as evidenced by: Based on a review of records and interview with the histotechnology laboratory supervisor, the laboratory failed to document the reactivity of the H&E (Hematoxylin & Eosin) stain each day of testing for one of ten days of patient testing. Findings include: (1) On 07/30/2025 at 09:45 am, the histotechnology laboratory supervisor stated the laboratory performed microscopic interpretations of histology (Mohs and skin biopsy) specimens that had been stained with H&E stain; (2) A review of test records for ten days of patient testing (microscopic interpretations) performed between 12/05/2024 and 07/10/2025 identified no evidence that the reactivity of the stain had been observed for acceptability for one (02/27/2025) of the ten days of patient testing reviewed; (3) The findings were reviewed with the histotechnology laboratory supervisor who stated on 07/30/2025 at 11:00 am, the reactivity of the stain was observed for acceptability on each patient slide but had not been documented as indicated above.</p>