

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 37D0962241	<b>(X3) Date Survey Completed</b> 12/07/2023
<b>Name of Provider or Supplier</b> Dermatology Surgery Center, Pc The	<b>Street Address, City, State</b> 1440 Terrace Drive, Tulsa, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The recertification survey was performed on 12/07/2023. The laboratory was found in compliance with a standard-level deficiency cited. The findings were reviewed with the office manager during an exit conference performed at the conclusion of the survey.
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with the histology technician, the laboratory failed to ensure expired supplies were not available for use. Findings include: (1) Observation of the laboratory on 12/07/2023 at 10:05 am, identified the following expired supplies that appeared to be available for use: (a) Tissue Marking Dyes: i. Avantik Red - lot #112131 with an expiration date of 12/31/2022 ii. Avantik Black - lot #112649 with an expiration date of 12/31/2022 iii. Avantik Green - lot #112722 with an expiration date of 12/31/2022 iv. Avantik Yellow - lot #111915 with an expiration date of 12/31/2022 v. CDI's Blue - lot #7275 with an expiration date of 10/2019 (b) One gallon bottle of Avantik Eosin Working Solution lot #245-04 with an expiration date of 10/04/2023 (2) Interview with the histology technician on 12/07/2023 at 11:35 am confirmed the expired supplies were available for use.</p>