

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D1014917	(X3) Date Survey Completed 10/27/2023
Name of Provider or Supplier Dermatology & Laser Center Of Oklahoma	Street Address, City, State 9306 S Toledo Court, Tulsa, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 11/27/2023. The laboratory was found in compliance with a standard-level deficiency cited. The findings were reviewed with the laboratory director and the testing person during an exit conference performed at the conclusion of the survey.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with testing person #2, the laboratory failed to ensure expired supplies were not available for use. Findings include: (1) Observation of the laboratory on 10/27/2023 at 09:17 am, identified the following expired supplies that appeared to be available for use: (a) Three bottles of CoverSeal-T Toluene Mounting Medium - lot # 96431 with an expiration date of 06/2022; (b) One bottle of CDI Tissue Marking Dye - Violet - lot # 20325 with an expiration date of 11/30/2022. (2) Interview with testing person #2 on 10/25/2023 at 10:02 am confirmed the expired supplies were available for use.</p>