

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D1054775	(X3) Date Survey Completed 06/17/2026
Name of Provider or Supplier Koweta Indian Health Facility	Street Address, City, State 31870 East State Highway 51, Coweta, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 06/17/2026. Standard-level deficiencies were cited.
D5211	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a) The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part. This STANDARD is not met as evidenced by: Based on a review of records and interview with technical consultant #4, the laboratory failed to review and evaluate proficiency testing results for one of four Chemistry Core Proficiency Testing events reviewed in 2025 and 2026. Findings include: (1) On 06/17/2026, a review of Chemistry Core Proficiency Testing records for four events (First 2025, Second 2025, Third 2025, and First 2026) identified the following failures with no evidence that corrective actions had been documented as performed: (a) First 2026 Event - The laboratory attained a score of 80% for Alanine Aminotransferase (ALT) (Sample CH-03); (2) Interview with technical consultant #4 on 06/17/2026 at 11:30 am confirmed corrective action had not been taken and documented for the failures.