

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  37D2018441	<b>(X3) Date Survey Completed</b>  02/14/2022
<b>Name of Provider or Supplier</b>  Tulsa Medical Laboratory, Llc	<b>Street Address, City, State</b>  1323 West 6th Avenue, Stillwater, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The recertification survey was performed on 02/14/2022. The findings were reviewed with the laboratory director at the conclusion of the survey. The laboratory was found in compliance with a standard-level deficiency cited.
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with the laboratory director, the laboratory failed to ensure that staining procedures were not performed with expired staining materials. Findings include: (1) On 02/14/2022 at 12:00 pm, the laboratory director stated the laboratory performed microscopic interpretations of histology specimens that had been stained with H&amp;E (Hematoxylin &amp; Eosin) stain; (2) On 02/14/2022 at 12:05 pm, the surveyor observed one coplin jar of 95% Alcohol (lot# 03418) with an expiration date of 01/31/2022 available for use; (3) The surveyor showed the coplin jar to the laboratory director, who stated to the surveyor on 02/14/2022 at 12:10 pm, the 95% Alcohol was not expired but that the label on the coplin jar had not been updated.</p>