

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D2022052	(X3) Date Survey Completed 04/14/2022
Name of Provider or Supplier Ascension St John Medical Center Emergency	Street Address, City, State 1923 S South Utica, Tulsa, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 04/14/2022. The findings were reviewed with the technical consultant and point of care technician at the conclusion of the survey. The laboratory was found in compliance with a standard-level deficiency cited.
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of records and interview with the technical consultant, the laboratory failed to review and evaluate proficiency testing results for one of 12 events. Findings include: (1) On 04/14/2022, the surveyor reviewed 2021 proficiency testing records and identified the following failure: (a) Second (AQI-B) Critical Care Aqueous Event (i) pO₂ - The laboratory failed the results for one of five samples (AQI-07). (2) A review of the records revealed there was no evidence proving the failure had been addressed; (3) The records were reviewed with the technical consultant. The technical consultant stated on 04/14/2022 at 02:35 pm corrective actions had not been taken.</p>