

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D2066492	(X3) Date Survey Completed 06/20/2019
Name of Provider or Supplier Saint Francis Lab - Sf Cancer Center	Street Address, City, State 11212 E 48th Street, Tulsa, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 06/20/19. The findings were reviewed with technical consultant/technical supervisor #1 at the conclusion of the survey. The laboratory was found out of compliance with the following CLIA regulations: 493.1413; D6033: Technical Consultant 493.1447; D6108: Technical Supervisor
D3031	<p>RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3)</p> <p>Analytic systems records. Retain quality control and patient test records (including instrument printouts, if applicable) and records documenting all analytic systems activities specified in 493.1252 through 493.1289 for at least 2 years.</p> <p>This STANDARD is not met as evidenced by: Based on a review of records and interview with technical consultant/technical supervisor #1, the laboratory failed to retain patient records for at least 2 years. Findings include: (1) At the beginning of the survey, technical consultant/technical supervisor #1 stated to the surveyor CBC (Complete Blood Count) testing was performed using the Cell-Dyn Emerald; (2) Later during the survey, the surveyor reviewed patient testing records with the following identified: (a) Records previous to 02/13/19 were not available (3) The surveyor ask technical consultant/technical supervisor #1 if patient records before 02/13/19 could be located; (4) Technical consultant/technical supervisor #1 stated to the surveyor the patient testing records before 02/13/19 for CBC testing could not be located.</p>
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p>

This STANDARD is not met as evidenced by:
 Based on a review of records, written policy and interview with technical consultant /technical supervisor #1, the laboratory failed to follow their written technical consultant and general supervisor competency policy based on the job responsibilities as listed in Subpart M. Findings include: (1) At the beginning of the survey, the surveyor reviewed personnel records for competency assessments performed during 2017, 2018 and 2019. There was no evidence competencies had been performed for the technical consultant and general supervisor, based on their job responsibilities; (2) The surveyor asked technical consultant/technical supervisor #1 if a written policy to evaluate the technical consultant and general supervisor based on job responsibilities was available. Technical consultant/technical supervisor #1 provided the policy for the surveyors review; (3) The surveyor reviewed the policy which required annual competencies be performed for the technical consultant and general supervisor based on job responsibilities. (4) The surveyor asked technical consultant/technical supervisor #1 if annual competencies based on job responsibilities had been performed during the review period. Technical consultant/technical supervisor #1stated annual competencies had not been performed.

D5211

EVALUATION OF PROFICIENCY TESTING PERFORMANCE
 CFR(s): 493.1236(a)

The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.

This STANDARD is not met as evidenced by:
 Based on a review of records and interview with technical consultant/technical supervisor #1 the laboratory failed to review and evaluate proficiency testing results. Findings include: (1) At the beginning of the survey, the surveyor reviewed 2017, 2018, and 2019 proficiency testing records. The following failure was identified, in which corrective action documentation could not be located: (a) Second 2018 Hematology Event (i) Blood Cell Identification - The laboratory failed the result for 1 of 5 samples, and attained a score of 80%. (2) The surveyor asked technical consultant /technical supervisor #1 if corrective action had been taken for the failure. After reviewing the records, technical consultant/technical supervisor #1 stated corrective action had not been taken for the failure.

D6033

TECHNICAL CONSULTANT-MODERATE COMPEXITY
 CFR(s): 493.1409

The laboratory must have a technical consultant who meets the qualification requirements of 493.1411 of this subpart and provides technical oversight in accordance with 493.1413 of this subpart.

This CONDITION is not met as evidenced by:
 Based on a review of records, and interview with technical consultant/technical supervisor #1, the technical consultant failed to provide technical oversight in accordance with 493.1413 of this subpart. Findings include: (1) The laboratory failed to ensure that individuals who performed the duties and responsibilities of the technical consultant, met the qualifications. Refer to D6035.

TECHNICAL CONSULTANT QUALIFICATIONS
CFR(s): 493.1411

(a) The technical consultant must be qualified and must possess a current license issued by the State in which the laboratory is located, if such licensing is required. (b) The technical consultant must-- (b)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (b)(1)(ii) Be certified in anatomic or clinical pathology, or both, by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (b)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (b)(2)(ii) Have at least one year of laboratory training or experience, or both in non-waived testing, in the designated specialty or subspecialty areas of service for which the technical consultant is responsible (for example, physicians certified either in hematology or hematology and medical oncology by the American Board of Internal Medicine are qualified to serve as the technical consultant in hematology); or (b)(3)(i) Hold an earned doctoral or master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (b)(3)(ii) Have at least one year of laboratory training or experience, or both in non-waived testing, in the designated specialty or subspecialty areas of service for which the technical consultant is responsible; or (b)(4)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (b)(4)(ii) Have at least 2 years of laboratory training or experience, or both in non-waived testing, in the designated specialty or subspecialty areas of service for which the technical consultant is responsible. Note: The technical consultant requirements for "laboratory training or experience, or both" in each specialty or subspecialty may be acquired concurrently in more than one of the specialties or subspecialties of service, excluding waived tests. For example, an individual who has a bachelor's degree in biology and additionally has documentation of 2 years of work experience performing tests of moderate complexity in all specialties and subspecialties of service, would be qualified as a technical consultant in a laboratory performing moderate complexity testing in all specialties and subspecialties of service.

This STANDARD is not met as evidenced by:

Based on a review of records and interview with technical consultant/technical supervisor #1, the laboratory failed to ensure that individuals who performed the duties and responsibilities of the technical consultant, met the qualifications. Findings include: (1) At the beginning of the survey, the surveyor reviewed records for 4 persons performing moderate complexity testing during 2017 and 2018. The records indicated the evaluations had been performed by an individual (testing person #1) with a degree from a foreign institution with no documentation of a United States (US) equivalency. Evaluations for 4 of 4 persons performing testing were not performed by a technical consultant: (a) Testing Person #2 - The 12/26/17 and 12/26/18 evaluations had been performed by testing person #1; (b) Testing Person #3 - The 07/03/17 and 07/19/18 evaluations had been performed by testing person #1; (c) Testing Person #4 - The 05/24/17 and 05/25/18 evaluations had been performed by testing person #1; (d) Testing Person #8 - The 09/03/17 and 09/03/18 evaluations had been performed by testing person #1. (2) The findings were reviewed with technical consultant/technical supervisor #1, who stated the evaluations had been performed by an individual with a foreign diploma without documentation of a US equivalency; (3)

The surveyor explained to technical consultant/technical supervisor #1 that all degrees from foreign institution must have an evaluation of their credentials to determine the equivalency of their education to an education obtained in the US.

D6108

LABORATORY TECHNICAL SUPERVISOR
CFR(s): 493.1447

The laboratory must have a technical supervisor who meets the qualification requirements of 493.1449 of this subpart and provides technical supervision in accordance with 493.1451 of this subpart.

This CONDITION is not met as evidenced by:
Based on a review of records and interview with technical consultant/technical supervisor #1, the technical supervisor failed to provide technical supervision in accordance with 493.1447 of this subpart. Findings include: (1) The technical supervisor failed to ensure the individual who performed the duties and responsibilities of the technical supervisor met the educational qualifications. Refer to D6111.

D6111

TECHNICAL SUPERVISOR QUALIFICATIONS
CFR(s): 493.1449

(a) The technical supervisor must possess a current license issued by the State in which the laboratory is located, if such licensing is required; and (b) The laboratory may perform anatomic and clinical laboratory procedures and tests in all specialties and subspecialties of services except histocompatibility and clinical cytogenetics services provided the individual functioning as the technical supervisor-- (b)(1) Is a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (b)(2) Is certified in both anatomic and clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or Possesses qualifications that are equivalent to those required for such certification. (c) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of bacteriology, the individual functioning as the technical supervisor must-- (c)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (c)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (c)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (c)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology; or (c)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (c)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology; or (c)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (c)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology

with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology; or (c)(5)(i) Have earned a bachelor's degree in a chemical, physical, or biological science or medical technology from an accredited institution; and (c)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology. (d) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of mycobacteriology, the individual functioning as the technical supervisor must-- (d)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (d)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (d)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (d)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (d)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (d)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (d)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology. (e) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of mycology, the individual functioning as the technical supervisor must-- (e)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (e)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (e)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (e)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology; or (e)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (e)(3)(ii) Have at least 1 year of laboratory training or experience, or both in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology; or (e)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (e)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity

testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology; or (e)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (e)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology. (f) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of parasitology, the individual functioning as the technical supervisor must-- (f)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (f)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (f)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (f)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; (f)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (f)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; or (f)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (f)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; or (f)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (f)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology. (g) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of virology, the individual functioning as the technical supervisor must-- (g)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (g)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (g)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (g)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology; or (g)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (g)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology; or (g)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (g)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity

testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology; or (g)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (g)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology. (h) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of diagnostic immunology, the individual functioning as the technical supervisor must-- (h)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (h)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (h)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (h)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology; or (h)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (h)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of diagnostic immunology; or (h)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (h)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology; or (h)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (h)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology. (i) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of chemistry, the individual functioning as the technical supervisor must-- (i)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (i)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (i)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (i)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry; or (i)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (i)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of chemistry; or (i)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (i)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry; or (i)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (i)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry. (j) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of hematology, the individual functioning as the technical supervisor must-- (j)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (j)(1)(ii) Be

certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (j)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (j)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing for the specialty of hematology (for example, physicians certified either in hematology or hematology and medical oncology by the American Board of Internal Medicine); or (j)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (j)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of hematology; or (j)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (j)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of hematology; or (j)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (j)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of hematology. (k)(1) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of cytology, the individual functioning as the technical supervisor must-- (k)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (k)(1)(ii) Meet one of the following requirements-- (k)(1)(ii)(A) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (k)(1)(ii)(B) Be certified by the American Society of Cytology to practice cytopathology or possess qualifications that are equivalent to those required for such certification; (l) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of histopathology, the individual functioning as the technical supervisor must-- (l)(1) Meet one of the following requirements: (l)(1)(i)(A) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (l)(1)(i)(B) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; (l)(1)(ii) An individual qualified under 493.1449(b) or paragraph (l)(1) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraph (b) or (l)(1)(i)(B) of this section, the responsibility for examination and interpretation of histopathology specimens. (l)(2) For tests in dermatopathology, meet one of the following requirements: (l)(2)(i)(A) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (l)(2)(i)(B) Meet one of the following requirements: (l)(2)(i)(B)(1) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(2)(i)(B)(2) Be certified in dermatopathology by the American Board of Dermatology and the American Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(2)(i)(B)(3) Be certified in dermatology by the American Board of Dermatology or possess qualifications that are equivalent to those required for such certification; or (l)(2)(ii) An individual qualified under 493.1449(b) or paragraph (l)(2)(i) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (l)(2)(i)(B) of this section, the

responsibility for examination and interpretation of dermatopathology specimens. (l) (3) For tests in ophthalmic pathology, meet one of the following requirements: (l)(3)(i) (A) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (l)(3)(i)(B) Must meet one of the following requirements: (l)(3)(i)(B)(1) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(3)(i)(B)(2) Be certified by the American Board of Ophthalmology or possess qualifications that are equivalent to those required for such certification and have successfully completed at least 1 year of formal post-residency fellowship training in ophthalmic pathology; or (l)(3)(ii) An individual qualified under 493.1449(b) or paragraph (1)(3)(i) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (1)(3)(i)(B) of this section, the responsibility for examination and interpretation of ophthalmic specimens; or (m) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of oral pathology, the individual functioning as the technical supervisor must meet one of the following requirements: (m)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (m)(1)(ii) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (m)(2) Be certified in oral pathology by the American Board of Oral Pathology or possess qualifications for such certification; or (m)(3) An individual qualified under 493.1449(b) or paragraph (m)(1) or (2) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (m)(1) or (2) of this section, the responsibility for examination and interpretation of oral pathology specimens. (n) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of radiobioassay, the individual functioning as the technical supervisor must-- (n)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (n)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (n)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (n)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay; or (n)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (n)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of radiobioassay; or (n)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (n)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay; or (n)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (n)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay. (o) If the laboratory performs tests in the specialty of histocompatibility, the individual functioning as the technical supervisor must either-- (o)(1)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (o)(1)(ii) Have training or experience that meets one of

the following requirements: (o)(1)(ii)(A) Have 4 years of laboratory training or experience, or both, within the specialty of histocompatibility; or (o)(1)(ii)(B)(1) Have 2 years of laboratory training or experience, or both, in the specialty of general immunology; and (o)(1)(ii)(B)(2) Have 2 years of laboratory training or experience, or both, in the specialty of histocompatibility; or (o)(2)(i) Have an earned doctoral degree in a biological or clinical laboratory science from an accredited institution; and (o)(2)(ii) Have training or experience that meets one of the following requirements: (o)(2)(ii)(A) Have 4 years of laboratory training or experience, or both, within the specialty of histocompatibility; or (o)(2)(ii)(B)(1) Have 2 years of laboratory training or experience, or both, in the specialty of general immunology; and (o)(2)(ii)(B)(2) Have 2 years of laboratory training or experience, or both, in the specialty of histocompatibility. (p) If the laboratory performs tests in the specialty of clinical cytogenetics, the individual functioning as the technical supervisor must-- (p)(1)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (p)(1)(ii) Have 4 years of training or experience, or both, in genetics, 2 of which have been in clinical cytogenetics; or (p)(2)(i) Hold an earned doctoral degree in a biological science, including biochemistry, or clinical laboratory science from an accredited institution; and (p)(2)(ii) Have 4 years of training or experience, or both, in genetics, 2 of which have been in clinical cytogenetics. (q) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of immunohematology, the individual functioning as the technical supervisor must-- (q)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (q)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (q)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (q)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing for the specialty of immunohematology. Note: The technical supervisor requirements for "laboratory training or experience, or both" in each specialty or subspecialty may be acquired concurrently in more than one of the specialties or subspecialties of service. For example, an individual, who has a doctoral degree in chemistry and additionally has documentation of 1 year of laboratory experience working concurrently in high complexity testing in the specialties of microbiology and chemistry and 6 months of that work experience included high complexity testing in bacteriology, mycology, and mycobacteriology, would qualify as the technical supervisor for the specialty of chemistry and the subspecialties of bacteriology, mycology, and mycobacteriology.

This STANDARD is not met as evidenced by:

Based on a review of records and interview with technical consultant/technical supervisor #1, the laboratory failed to ensure that individuals who performed the duties and responsibilities of the technical supervisor, met the qualifications. Findings include: (1) At the beginning of the survey, the surveyor reviewed records for 4 persons performing high complexity testing during 2017 and 2018. The records indicated the evaluations had been performed by an individual (testing person #1) with a degree from a foreign institution with no documentation of a United States (US) equivalency. Evaluations for 4 of 4 persons performing testing were not performed by a technical supervisor: (a) Testing Person #2 - The 12/26/17 and 12/26/18 evaluations had been performed by testing person #1; (b) Testing Person #3 - The

07/03//17 and 07/19/18 evaluations had been performed by testing person #1; (c) Testing Person #4 - The 05/24/17 and 05/25/18 evaluations had been performed by testing person #1; (d) Testing Person #8 - The 09/03/17 and 09/03/18 evaluations had been performed by testing person #1. (2) The findings were reviewed with technical consultant/technical supervisor #1, who stated the evaluations had been performed by an individual with a foreign diploma without documentation of a US equivalency; (3) The surveyor explained to technical consultant/technical supervisor #1 that all degrees from foreign institution must have an evaluation of their credentials to determine the equivalency of their education to an en education obtained in the US.