

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D2082356	(X3) Date Survey Completed 10/01/2020
Name of Provider or Supplier Xyz Medical, Llc	Street Address, City, State 104 W Ohio St, Ste 200, Coalgate, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The initial survey was performed on 09/30/2020 and 10/01/2020. The findings were reviewed with the general supervisor at the conclusion of the survey. The laboratory was found out of compliance with the following CLIA regulations: 493.1447; D6108: Technical Supervisor
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on a review of records and interview with the general supervisor, the laboratory failed to have a written competency policy for the technical supervisor and general supervisor based on the job responsibilities as listed in Subpart M. Findings include: (1) On 09/30/2020, surveyor #2 reviewed personnel records for competency assessments performed during 2019 through 09/30/2020. There was no evidence competencies had been performed for the technical supervisor and general supervisor based on their job responsibilities; (2) On 10/01/2020, surveyor #2 asked the general supervisor if a written policy to evaluate the positions based on job responsibilities was available and if competencies had been performed during the review period. The general supervisor stated on 10/01/2020 at 1:15 pm a policy to evaluate the clinical consultant based on job responsibilities had not been written; and competencies had not been performed.</p>
D5311	<p>SPECIMEN SUBMISSION, HANDLING, AND REFERRAL CFR(s): 493.1242(a)</p> <p>The laboratory must establish and follow written policies and procedures for each of</p>

the following, if applicable: (1) Patient preparation. (2) Specimen collection. (3) Specimen labeling, including patient name or unique patient identifier and, when appropriate, specimen source. (4) Specimen storage and preservation. (5) Conditions for specimen transportation. (6) Specimen processing. (7) Specimen acceptability and rejection. (8) Specimen referral.

This STANDARD is not met as evidenced by:
Based on a review of the procedure manual and interview with the general supervisor, the laboratory failed to have a written procedure for specimen submission, handling and processing of turbid specimens for urine drug confirmatory testing. Findings include: (1) On 09/30/2020, surveyor #1 reviewed the sections of the procedure manual titled, "LCMS" and "General Laboratory Operations". A specimen submission, handling and processing procedure could not be located for how to submission, handling, and processing of turbid specimens received in the laboratory for testing; (2) Surveyor #1 explained to the general supervisor, that a specimen submission, handling and processing procedure could not be located. The general supervisor stated on 09/30/2020 at 03:45 pm that a procedure had not been written.

D5401

PROCEDURE MANUAL
CFR(s): 493.1251(a)

A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.

This STANDARD is not met as evidenced by:
Based on a review of a written policy and procedure manual and interview with the general supervisor, the laboratory failed to have written procedures that explained the current practices and procedures of stock solution preparation being performed in the laboratory for urine drug confirmatory testing. Findings include: (1) On 09/30/2020 at 01:00 pm, the general supervisor stated urine drug confirmatory testing was performed using the AB Sciex 4500 analyzer; (2) Surveyor #2 reviewed the manual titled, "LCMS Guide Standard Operation Procedure: XYZ Panel-55 section titled, "Stock Solution Preparation", stated the following: (a) "...Total volume of combined stocks at the step is 10.34 mL of solution in Methanol." (b) "Complete to 20 mL by adding 9.66mL of Methanol into the tube." (c) "Vortex to mix." (3) On 10/01/2020 surveyor #2 was not able to locate a vortex in the laboratory. Surveyor #2 asked the general supervisor if the solution was vortexed. On 10/01/2020 at 10:10 am the general supervisor stated the solution was inverted not vortexed as indicated above.

D5403

PROCEDURE MANUAL
CFR(s): 493.1251(b)

The procedure manual must include the following when applicable to the test procedure: (1) Requirements for patient preparation; specimen collection, labeling, storage, preservation, transportation, processing, and referral; and criteria for specimen acceptability and rejection as described in 493.1242. (2) Microscopic examination, including the detection of inadequately prepared slides. (3) Step-by-step performance of the procedure, including test calculations and interpretation of results. (4) Preparation of slides, solutions, calibrators, controls, reagents, stains, and other

materials used in testing. (5) Calibration and calibration verification procedures. (6) The reportable range for test results for the test system as established or verified in 493.1253. (7) Control procedures. (8) Corrective action to take when calibration or control results fail to meet the laboratory's criteria for acceptability. (9) Limitations in the test methodology, including interfering substances. (10) Reference intervals (normal values). (11) Imminently life-threatening test results, or panic or alert values. (12) Pertinent literature references. (13) The laboratory's system for entering results in the patient record and reporting patient results including, when appropriate, the protocol for reporting imminently life threatening results, or panic, or alert values. (14) Description of the course of action to take if a test system becomes inoperable.

This STANDARD is not met as evidenced by:
 Based on a review of the policy and procedure manual and interview with the general supervisor, the laboratory failed to have a written procedure to validate a column change for the LCMS analyzer. Findings include: (1) On 09/30/2020, surveyor #1 reviewed the sections of the procedure manual titled, "LCMS" and "General Laboratory Operations". A procedure could not be located to validate a column change; (2) Surveyor #1 explained to the general supervisor, that the laboratory must document when a column change is performed and the laboratory's procedure on how to ensure the column is functioning properly. The general supervisor stated on 09/30/2020 at 03:50 pm that a procedure had not been written.

D5421

ESTABLISHMENT AND VERIFICATION OF PERFORMANCE
 CFR(s): 493.1253(b)(1)

Each laboratory that introduces an unmodified, FDA-cleared or approved test system must do the following before reporting patient test results: (1)(i) Demonstrate that it can obtain performance specifications comparable to those established by the manufacturer for the following performance characteristics: (1)(i)(A) Accuracy. (1)(i)(B) Precision. (1)(i)(C) Reportable range of test results for the test system. (1)(ii) Verify that the manufacturer's reference intervals (normal values) are appropriate for the laboratory's patient population.

This STANDARD is not met as evidenced by:
 Based on a review of records and interview with the general supervisor, the laboratory failed to verify the acceptable performance of the laboratory computer system. Findings include: (1) On 09/30/2021 at 01:30 pm, the general supervisor stated to the surveyors the laboratory began patient urine drug confirmatory testing on 05/08/2020 using the AB Sciex 4500 analyzer, which included installing the Lab Daq laboratory information system (LIS); (2) On 10/01/2020, the surveyors reviewed the validation records for the Lab Daq LIS and could not locate documentation to prove the data had been reviewed by the laboratory before putting into use for patient testing; (3) The surveyors reviewed the records with the general supervisor and asked if there was evidence the laboratory reviewed the LIS validation before it was integrated into routine operation. The general supervisor stated to the surveyors on 10/01/2020 at 10:30 am, there was no documentation to prove the LIS validation had been reviewed and accepted prior to being used for patient testing.

D5435

MAINTENANCE AND FUNCTION CHECKS
 CFR(s): 493.1254(b)(2)

For equipment, instruments, or test systems developed in-house, commercially available and modified by the laboratory, or maintenance and function check protocols are not provided by the manufacturer, the laboratory must: (i) Define a function check protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting. (ii) Perform and document the function checks, including background or baseline checks, specified in paragraph (b)(2)(i) of this section. Function checks must be within the laboratory's established limits before patient testing is conducted.

This STANDARD is not met as evidenced by:

Based on a review of records, written policies and procedures, and interview with the general supervisor, the laboratory failed to define a function check protocol to ensure the centrifuge and pipettes were functioning properly. Findings include: (1) On 09/30/2020 on 01:00 pm, the general supervisor stated the following to the surveyors: (a) Urine drug confirmatory testing was performed using the AB Sciex 4500 analyzer; (b) Specimens were processed in the Thermofisher Sorvall ST8 3700 rpm (revolutions per minute) for 15 minutes and 1800 rpm for 10 minutes for turbid specimens; (c) The following pipettes were used in the stock solution preparation: (i) Eppendorf Repeater M4 (serial number F10463C) (ii) Labsystems 4510 (serial number E69549) (iii) ThermoScientific Finnipipette F1 (serial number LH05729) (iv) ThermoScientific Finnipipette F1 (serial number LH13455) (v) ThermoScientific Finnipipette F1 (serial number LH09702) (vi) ThermoScientific Finnipipette F1 (serial number LH37968) (2) Surveyor #2 reviewed the laboratory's written policy and procedure manuals titled, "LCMS" and "General Laboratory Operations" and was not able to locate a procedure describing how to check the following: (a) Centrifuge speed, time, and the frequency the checks were to be performed; (b) Accuracy of the pipettes and the frequency the checks were to be performed. (3) Surveyor #2 asked the general supervisor if centrifuge and pipette function check protocols were written. The general supervisor stated on 09/30/2020 at 03:45 pm a function check protocol for the centrifuge and pipettes had not been written.

D5805

TEST REPORT

CFR(s): 493.1291(c)

The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.

This STANDARD is not met as evidenced by:

Based on a review of patient reports and interview with the general supervisor, the laboratory failed to ensure reports included accurate test report dates for 19 of 19 patient reports. Findings include: (1) On 09/30/2020 at 01:00 pm, the general supervisor stated to the surveyors the laboratory began performing urine drug confirmatory testing using the AB Sciex 4500 analyzer on 05/08/2020; (2) On 10/01/2020, surveyor #1 reviewed 19 patient reports for urine drug confirmatory testing collected from 05/05/2020 through 06/16/2020. For 19 of 19 reports, the test date was

documented as 09/18/2020; (3) Surveyor #1 reviewed the reports with the general supervisor and asked if 09/18/2020 was the correct report date for the 19 reports. The general supervisor looked in the analyzers' memory for each report and obtained the actual report dates for the surveyor, which correlated to the collection and received dates for each patient. The general supervisor stated on 10/01/2020 at 11:00 am, the test dates on the patient reports were not reflective of the actual test dates because the LIS server was updated on 09/18/2020, which resulted in the test dates changing to that date.

D6108

LABORATORY TECHNICAL SUPERVISOR
CFR(s): 493.1447

The laboratory must have a technical supervisor who meets the qualification requirements of 493.1449 of this subpart and provides technical supervision in accordance with 493.1451 of this subpart.

This CONDITION is not met as evidenced by:

Based on a review of records and interview with the general supervisor, the technical supervisor failed to provide technical oversight in accordance with 493.1413 of this subpart for high complexity testing. Findings include: (1) The technical supervisor failed to ensure that individuals who performed the duties and responsibilities of a technical supervisor met the required educational qualifications. Refer to D6111.

D6111

TECHNICAL SUPERVISOR QUALIFICATIONS
CFR(s): 493.1449

(a) The technical supervisor must possess a current license issued by the State in which the laboratory is located, if such licensing is required; and (b) The laboratory may perform anatomic and clinical laboratory procedures and tests in all specialties and subspecialties of services except histocompatibility and clinical cytogenetics services provided the individual functioning as the technical supervisor-- (b)(1) Is a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (b)(2) Is certified in both anatomic and clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or Possesses qualifications that are equivalent to those required for such certification. (c) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of bacteriology, the individual functioning as the technical supervisor must-- (c)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (c)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (c)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (c)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology; or (c)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (c)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the

subspecialty of bacteriology; or (c)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (c)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology; or (c)(5)(i) Have earned a bachelor's degree in a chemical, physical, or biological science or medical technology from an accredited institution; and (c)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology. (d) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of mycobacteriology, the individual functioning as the technical supervisor must-- (d)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (d)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (d)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (d)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (d)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (d)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (d)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology. (e) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of mycology, the individual functioning as the technical supervisor must-- (e)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (e)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (e)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (e)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology; or (e)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (e)(3)(ii) Have at least 1 year of laboratory training or experience, or both in high complexity testing within the specialty of microbiology with a minimum of 6 months

experience in high complexity testing within the subspecialty of mycology; or (e)(4) (i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (e)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology; or (e)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (e)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology. (f) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of parasitology, the individual functioning as the technical supervisor must-- (f)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (f)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (f)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (f)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; (f)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (f)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; or (f)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (f)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; or (f)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (f)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology. (g) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of virology, the individual functioning as the technical supervisor must-- (g)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (g)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (g)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (g)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology; or (g)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (g)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months

experience in high complexity testing within the subspecialty of virology; or (g)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (g)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology; or (g)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (g)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology. (h) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of diagnostic immunology, the individual functioning as the technical supervisor must-- (h)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (h)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (h)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (h)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology; or (h)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (h)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of diagnostic immunology; or (h)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (h)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology; or (h)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (h)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology. (i) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of chemistry, the individual functioning as the technical supervisor must-- (i)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (i)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (i)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (i)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry; or (i)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (i)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of chemistry; or (i)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (i)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry; or (i)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (i)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry. (j) If the requirements

of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of hematology, the individual functioning as the technical supervisor must-- (j)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (j)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (j)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (j)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing for the specialty of hematology (for example, physicians certified either in hematology or hematology and medical oncology by the American Board of Internal Medicine); or (j)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (j)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of hematology; or (j)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (j)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of hematology; or (j)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (j)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of hematology. (k)(1) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of cytology, the individual functioning as the technical supervisor must-- (k)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (k)(1)(ii) Meet one of the following requirements-- (k)(1)(ii)(A) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (k)(1)(ii)(B) Be certified by the American Society of Cytology to practice cytopathology or possess qualifications that are equivalent to those required for such certification; (l) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of histopathology, the individual functioning as the technical supervisor must-- (l)(1) Meet one of the following requirements: (l)(1)(i)(A) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (l)(1)(i)(B) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; (l)(1)(ii) An individual qualified under 493.1449(b) or paragraph (l)(1) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraph (b) or (l)(1)(i)(B) of this section, the responsibility for examination and interpretation of histopathology specimens. (l)(2) For tests in dermatopathology, meet one of the following requirements: (l)(2)(i)(A) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (l)(2)(i)(B) Meet one of the following requirements: (l)(2)(i)(B)(1) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(2)(i)(B)(2) Be certified in dermatopathology by the American Board of Dermatology and the American Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(2)(i)(B)(3) Be certified in dermatology by the American Board of Dermatology or possess

qualifications that are equivalent to those required for such certification; or (l)(2)(ii) An individual qualified under 493.1449(b) or paragraph (l)(2)(i) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (l)(2)(i)(B) of this section, the responsibility for examination and interpretation of dermatopathology specimens. (l)(3) For tests in ophthalmic pathology, meet one of the following requirements: (l)(3)(i) (A) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (l)(3)(i)(B) Must meet one of the following requirements: (l)(3)(i)(B)(1) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(3)(i)(B)(2) Be certified by the American Board of Ophthalmology or possess qualifications that are equivalent to those required for such certification and have successfully completed at least 1 year of formal post-residency fellowship training in ophthalmic pathology; or (l)(3)(ii) An individual qualified under 493.1449(b) or paragraph (l)(3)(i) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (l)(3)(i)(B) of this section, the responsibility for examination and interpretation of ophthalmic specimens; or (m) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of oral pathology, the individual functioning as the technical supervisor must meet one of the following requirements: (m)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (m)(1)(ii) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (m)(2) Be certified in oral pathology by the American Board of Oral Pathology or possess qualifications for such certification; or (m)(3) An individual qualified under 493.1449(b) or paragraph (m)(1) or (2) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (m)(1) or (2) of this section, the responsibility for examination and interpretation of oral pathology specimens. (n) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of radiobioassay, the individual functioning as the technical supervisor must-- (n)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (n)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (n)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (n)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay; or (n)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (n)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of radiobioassay; or (n)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (n)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay; or (n)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (n)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay. (o) If the laboratory performs tests in the specialty of

histocompatibility, the individual functioning as the technical supervisor must either-- (o)(1)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (o)(1)(ii) Have training or experience that meets one of the following requirements: (o)(1)(ii)(A) Have 4 years of laboratory training or experience, or both, within the specialty of histocompatibility; or (o)(1)(ii)(B)(1) Have 2 years of laboratory training or experience, or both, in the specialty of general immunology; and (o)(1)(ii)(B)(2) Have 2 years of laboratory training or experience, or both, in the specialty of histocompatibility; or (o)(2)(i) Have an earned doctoral degree in a biological or clinical laboratory science from an accredited institution; and (o)(2)(ii) Have training or experience that meets one of the following requirements: (o)(2)(ii)(A) Have 4 years of laboratory training or experience, or both, within the specialty of histocompatibility; or (o)(2)(ii)(B)(1) Have 2 years of laboratory training or experience, or both, in the specialty of general immunology; and (o)(2)(ii)(B)(2) Have 2 years of laboratory training or experience, or both, in the specialty of histocompatibility. (p) If the laboratory performs tests in the specialty of clinical cytogenetics, the individual functioning as the technical supervisor must-- (p)(1)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (p)(1)(ii) Have 4 years of training or experience, or both, in genetics, 2 of which have been in clinical cytogenetics; or (p)(2)(i) Hold an earned doctoral degree in a biological science, including biochemistry, or clinical laboratory science from an accredited institution; and (p)(2)(ii) Have 4 years of training or experience, or both, in genetics, 2 of which have been in clinical cytogenetics. (q) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of immunohematology, the individual functioning as the technical supervisor must-- (q)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (q)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (q)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (q)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing for the specialty of immunohematology. Note: The technical supervisor requirements for "laboratory training or experience, or both" in each specialty or subspecialty may be acquired concurrently in more than one of the specialties or subspecialties of service. For example, an individual, who has a doctoral degree in chemistry and additionally has documentation of 1 year of laboratory experience working concurrently in high complexity testing in the specialties of microbiology and chemistry and 6 months of that work experience included high complexity testing in bacteriology, mycology, and mycobacteriology, would qualify as the technical supervisor for the specialty of chemistry and the subspecialties of bacteriology, mycology, and mycobacteriology.

This STANDARD is not met as evidenced by:

Based on a review of records and interview with the general supervisor, the technical supervisor failed to ensure that individuals who performed the duties and responsibilities of the technical supervisor, met the qualifications for 1 of 1 testing person. Findings include: (1) On 09/30/2020, surveyor #2 reviewed records for 1 person performing high complexity since the laboratory began patient testing on 05/08/2020. The records showed the semi-annual evaluation for 1 of 1 person had been

performed by an individual who did not meet the regulatory qualification requirements of the technical supervisor: (a) Testing Person #2 (i) The 08/28/2020 semi-annual evaluation had been performed by testing person #1 (this person had earned a bachelor of science in biology but did not meet the 4 years of experience requirement). (2) Surveyor #2 explained to the general supervisor on 09/30/2020 at 04:00 pm that all components of the competency evaluations must be performed by a person who qualifies as a technical supervisor (an individual with a minimum of a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution, and at least 4 years of laboratory training or experience in the designated specialty or subspecialty areas of service). The general supervisor stated on 09/30/2020 at 04:00 pm, the above evaluation had been performed by an individual who did not meet the qualifications.

D6127

TECHNICAL SUPERVISOR RESPONSIBILITIES

CFR(s): 493.1451(b)(9)

The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least semiannually during the first year the individual tests patient specimens.

This STANDARD is not met as evidenced by:
Based on a review of records and interview with the general supervisor, the technical supervisor failed to ensure persons were evaluated at least semiannually during the first year of performing high complexity testing for 1 of 1 testing person. Findings include: (1) On 09/30/2020, surveyor #2 reviewed personnel records. The following was identified for 1 of 1 testing person: (a) Testing Person #1 - This person completed initial training 01/09/2019 and annual training 02/10/2020. There was no evidence that a semiannual evaluation had been performed. (2) On 10/01/2020, surveyor #2 reviewed the findings with the general supervisor who stated on 10/01/2020 at 01:20 pm a semiannual evaluation had not been performed.