

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D2123513	(X3) Date Survey Completed 02/05/2026
Name of Provider or Supplier Nga Lab Services Llc	Street Address, City, State 1515 N Porter Ave, Suite 200, Norman, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 02/05/2026. Standard-level deficiencies were cited.
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>(a) A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of policy, direct observation, and interview with the laboratory director, the laboratory failed to follow their policy for schiff stain storage for one of one bottle observed. Findings include: (1) On 02/05/2026 at 09:45 am, the laboratory director stated the laboratory performs alcian blue/periodic acid schiff (PAS) staining; (2) A review of the policy titled "Alcian Blue /PAS Protocol", stated the following: (a) "Schiff's is to be stored in the refrigerator after opening"; (3) Observation of the laboratory revealed an open bottle of schiff's stain, stored at room temperature, which stated, "refrigerate after opening". (4) Interview with the laboratory director on 02/05/2026 at 09:45 am, confirmed the laboratory failed to follow their policy for storage of the schiff stain after opening.</p>