

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 37D2137232 | (X3) Date Survey Completed 04/04/2023 |
| Name of Provider or Supplier Epiphany Dermatology Of Oklahoma, Llc | Street Address, City, State 3111 Azalea Park Drive, Muskogee, OK | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D0000 | The recertification survey was performed on 04/04/2023. The laboratory was found in compliance with a standard-level deficiency cited. The findings were reviewed with the office manager at the conclusion of the survey. |
| D5407 | <p>PROCEDURE MANUAL CFR(s): 493.1251(d)</p> <p>Procedures and changes in procedures must be approved, signed, and dated by the current laboratory director before use.</p> <p>This STANDARD is not met as evidenced by: Based on a review of a written procedure and interview with the office manager, the laboratory failed to ensure a written procedure had been approved, signed, and dated by the laboratory director. Findings include: (1) On 04/04/2023 at 11:50 am, the office manager stated KOH (Potassium Hydroxide) testing was performed as a PPM (Provider Performed Microscopy) procedure; (2) A review of the procedure titled, "KOH Protocol" identified no indication it had been approved, signed, and dated by the laboratory director; (3) The findings were reviewed with the office manager who stated on 04/04/2023 at 01:50 pm, the procedure had not been signed and dated by the laboratory director.</p> |