Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D2145906	(X3) Date Survey Completed 12/18/2024
Name of Provider or Supplier Classen Urgent Care Moore	**Street Address, City, State** 1015 Sw 4th Street, Moore, OK	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 12/18/2024. The laboratory was found in compliance with standard-level deficiencies cited. The findings were reviewed with the technical consultant at the conclusion of the survey.
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on a review of records, observation, and interview with the technical consultant, the laboratory failed to ensure eight of eight Quantiferon (QFT) tubes had not exceeded their expiration date and and were available for use. Findings include: (1) On 12/18/2024 at 10:00 am, the techincal consultant stated that the laboratory collected blood for quantiferon testing that was sent to a reference lab for testing; (2) Observation of the laboratory on 12/18/2024 at 10:00 am identified eight QFT tubes had exceeded thier expiration dates: (a) Two yellow top QFT tubes, lot #1083244 with an expiration date of 3/31/2024; (b) Two grey top QFT tubes, lot #1083242 with an expiration date of 3/31/2024; (c) Two purple top QFT tubes, lot #1083245 with an expiration date of 04/30/2024; (d) Two green top QFT tubes, lot #1083243 with an expiration date of 3/31/2024. (3) Interview with the technical consultant on 12/18/2024 at 10:00 am confirmed the quantiferon tubes were expired and available for use.