

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 37D2147408	<b>(X3) Date Survey Completed</b> 06/28/2023
<b>Name of Provider or Supplier</b> Advanced Pathology	<b>Street Address, City, State</b> 4208 Sw Lee Blvd, Lawton, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The recertification survey was performed on 06/28/2023. The laboratory was found in compliance with a standard-level deficiency cited. The findings were reviewed with the laboratory director at the conclusion of the survey.
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with the histology supervisor, the laboratory failed to ensure expired supplies were not available for use. Findings include: (1) Observation of the laboratory on 06/28/2023 at 10:27 am, identified the following expired reagent that appeared to be available for use: (a) Mercedes Scientific Silver Nitrate Solution (0.2%) - lot #61961 with an expiration date of 12/2022. (2) Interview with the histology supervisor on 06/28/202 at 10:30 a.m. confirmed the Silver Nitrate Solution was used for special stains and was available for use.</p>