

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 37D2278351	(X3) Date Survey Completed 06/23/2026
Name of Provider or Supplier Premier Plus Urgent Care	Street Address, City, State 12000 Charleston Way, Yukon, OK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The recertification survey was performed on 06/23/2026. Standard-level deficiencies were cited.
D2010	TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(2) (b)(2) The laboratory must test samples the same number of times that it routinely tests patient samples. This STANDARD is not met as evidenced by: Based on a review of records and interview with testing person #1, the laboratory failed to test proficiency testing samples the same number of times that patient samples were tested for two of four American Association of Bioanalysts Medical Laboratory Evaluation (AAB-MLE) proficiency testing events reviewed. Findings include: (1) On 06/23/2026 at 11:30 am testing person #1 stated CBC (Complete Blood Count) testing was performed using the Sysmex poch-100I analyzer; (2) A review of AAB-MLE proficiency testing records for four events (First 2025, Second 2025, Third 2025, First 2026) identified specimens had been tested in duplicate as follows: (a) First 2025 Event - Five of five samples were tested in duplicate for CBC testing; (i) Sample 01 was tested on 02/20/2025 at 10:00 am and 02/20/2025 at 10:26 am; (ii) Sample 02 was tested on 02/20/2025 at 10:31 am and 02/20/2025 at 10:42 am; (iii) Sample 03 was tested on 02/20/2025 at 11:53 am and 02/20/2025 at 12:13 pm; (iv) Sample 04 was tested on 02/20/2025 at 12:51 pm and 02/20/2025 at 01:13 pm; (v) Sample 05 was tested on 02/20/2025 at 01:41 pm and 02/20/2025 at 02:12 pm. (b) Second 2025 Event - Five of five samples were tested in duplicate for CBC testing; (i) Sample 06 was tested on 05/20/2025 at 01:23 pm and 05/20/2025 at 01:42 pm. (ii) Sample 07 was tested on 05/20/2025 at 01:05 pm and 05/20/2025 at 01:15 pm; (iii) Sample 08 was tested on 05/20/2025 at 12:57 pm and 05/20/2025 at 01:03 pm; (iv) Sample 09 was tested on 05/20/2025 at 12:50 pm and 05/20/2025 at 12:47 pm; (v) Sample 10 was tested on 05/20/2025 at 12:36 pm and 05/20/2025 at 12:43 pm. (3)

The records were reviewed with testing person #1 who stated on 06/23/20269 at 11:30 am, the samples had been tested in duplicate as stated above and patient samples were not routinely tested in this manner.