

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  38D0064851	<b>(X3) Date Survey Completed</b>  08/02/2022
<b>Name of Provider or Supplier</b>  University Health Services Laboratory	<b>Street Address, City, State</b>  1590 E 13th Ave, Eugene, OR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p><b>SUCCESSFUL PARTICIPATION</b> CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Proficiency Testing (PT) desk review of the American Proficiency Institute (API) proficiency testing showed the laboratory had two consecutive unsuccessful participation 1st and 2nd event 2022 in Endocrinology for the analyte Human Chorionic Gonadotropin (HCG). Refer to D2099 and D2107.</p>
<b>D2099</b>	<p><b>ENDOCRINOLOGY</b> CFR(s): 493.843(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory</p>

performance.

This STANDARD is not met as evidenced by:  
Proficiency Testing (PT) desk review of the American Proficiency Institute (API) proficiency testing showed the laboratory failed to attain an overall testing event score of 80 percent for two consecutive events in the specialty Endocrinology. Findings include: 1. 1st event 2022 Endocrinology = 60% 2. 2nd event 2022 Endocrinology = 60%

**D2107**

**ENDOCRINOLOGY**

CFR(s): 493.843(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:  
Proficiency Testing (PT) desk review of the American Proficiency Institute (API) proficiency testing showed the laboratory has had unsuccessful performance for two consecutive testing events for the analyte Human Chorionic Gonadotropin (HCG) in the specialty Endocrinology. Findings includes: 1. 1st Event 2022, HCG = 60% 2. 2nd Event 2022, HCG = 60%