

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 38D0628271	(X3) Date Survey Completed 06/26/2018
Name of Provider or Supplier Bend Dermatology Clinic Lab Llc	Street Address, City, State 2747 Ne Connors Avenue, Bend, OR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6120	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(7)(8)</p> <p>(7) The technical supervisor is responsible for identifying training needs and assuring that each individual performing tests receives regular in-service training and education appropriate for the type and complexity of the laboratory services performed; (8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Review of competency records and discussion with the staff the Laboratory Director who also served as the Technical Supervisor failed to assess and document competency testing of the testing personnels. Findings include. 1. 4 out of 4 testing pesonnel who are dermatologist do not have documentations of annual competency evaluation for the year 2016, 2017 and 2018. Last competency documentations was 04 /14/2015. 2. This fnding was confirmed by the Histology staff on 06/26/2018 at 11: 30AM.</p>