

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 38D0628651	(X3) Date Survey Completed 10/09/2018
Name of Provider or Supplier Blue Mountain Hospital	Street Address, City, State 170 Ford Rd, John Day, OR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Review of the College of American Pathologists (Excel) (CAP) proficiency testing (PT) shows laboratory had unsuccessful participation for in Immunohematology. Refer to D2182.
D2182	ANTIBODY IDENTIFICATION CFR(s): 493.865(a) Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.

This STANDARD is not met as evidenced by:
Review of the College of American Pathologist (CAP) proficiency testing (PT)
reveals that the laboratory has unsatisfactory performance in antibody identification.
Finding include: 1. 2nd Event of 2018 - Antibody Identification - 0%.