

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 38D0671650	(X3) Date Survey Completed 10/09/2018
Name of Provider or Supplier Good Shepherd Health Care System	Street Address, City, State 610 Nw 11th St, Hermiston, OR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Record review of the American Association of Bioanalyst (AAB) proficiency testing reveals that your laboratory has unsuccessful proficiency testing for Diagnostic Immunology. Refer to D2074.</p>
D2074	<p>SYPHILIS SEROLOGY CFR(s): 493.835(e)</p> <p>Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is</p>

unsuccessful performance.

This STANDARD is not met as evidenced by:

Review of the American Association of Bioanalyst (AAB) proficiency testing show unsatisfactory performance of two consecutive testing events. Findings include: 1. 1st Event 2018 - Syphilis Serology - 20% 2. 2nd Event 2018 - Syphilis Serology - 0%