

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 38D0716183	(X3) Date Survey Completed 10/04/2021
Name of Provider or Supplier St Charles Prineville Laboratory	Street Address, City, State 384 Se Combs Flat Road, Prineville, OR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Review of the College of American Pathologists (CAP) proficiency testing (PT) showed the laboratory had consecutive unsuccessful participation for the 1st and 2nd event 2021 for Antibody Identification in the specialty Immunohematology . Refer to D2191.
D2191	ANTIBODY IDENTIFICATION CFR(s): 493.865(f) Failure to achieve an overall testing event score of satisfactory for two consecutive

testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Review of the College of American Pathologist (CAP) proficiency testing (PT) showed that the laboratory has unsatisfactory performance for antibody identification in the specialty Immunohematology. Finding include: 1. 1st Event 2021 - Antibody Identification - 0%. 2. 2nd Event 2021- Antibody Identification - 0%.