

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 38D0883097	(X3) Date Survey Completed 12/03/2020
Name of Provider or Supplier Columbia View Family Health Center	Street Address, City, State 2800 Sw 257th Drive, Troutdale, OR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Proficiency Testing (PT) desk review of the American Association of Bioanalysts (AAB) proficiency testing (PT) shows the laboratory had unsuccessful participation for the Analytes Creatine kinase and Uric Acid in Routine Chemistry. Refer to D2096.</p>
D2096	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p> <p>Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is</p>

unsuccessful performance.

This STANDARD is not met as evidenced by:

Record review of proficiency testing from the American Association of Bioanalysts (AAB) reveals that your laboratory has unsuccessful proficiency testing in the following analytes in routine chemistry. The findings include: 1. 2nd Event 2020 - Creatine kinase - 0% 3rd Event 2020 - Creatine kinase - 0% 2. 2nd Event 2020 - Uric Acid - 0% 3rd Event 2020 - Uric Acid - 0%