

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  38D1052681	<b>(X3) Date Survey Completed</b>  02/17/2026
<b>Name of Provider or Supplier</b>  Laboratory Corporation Of America	<b>Street Address, City, State</b>  255 Lancaster Dr Ne, Salem, OR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5445</b>	<p>CONTROL PROCEDURES CFR(s): 493.1256(d)(1)(2)(g)</p> <p>(d) Unless CMS Approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing, the laboratory must-- (d)(1) Perform control procedures as defined in this section unless otherwise specified in the additional specialty and subspecialty requirements at 493.1261 through 493.1278. (d)(2) For each test system, perform control procedures using the number and frequency specified by the manufacturer or established by the laboratory when they meet or exceed the requirements in paragraph (d)(3) of this section. (d)(3) At least once each day patient specimens are assayed or examined perform the following for:</p> <p>This STANDARD is not met as evidenced by: Based on review of the Quality Control (QC) records for January 2026 and interview with the Laboratory Director (LD), the laboratory failed to ensure that QC was performed by all persons performing urinalysis on patient specimens. Findings include: 1. Upon review of QC records for urinalysis for January 2026, using Bio-Rad controls #1 &amp; #2, lot numbers 79981 exp. 08/01/2026 and 79982 exp. 08/01/2026, it was revealed that one (1) of four (4) testing personnel (TP #1) performed the QC on the urine dipsticks twenty (20) days out of twenty six (26) working days in January 2026. 2. TP #2 performed QC on the urine dipsticks six (6) days out of twenty six (26) days in January 2026.. 3. TP #3 performed no QC on the urine dipsticks in January 2026. 4. TP #4 performed no QC on the urine dipsticks in January 2026. 5. Interview with the Laboratory Director (LD) at 2:30 pm confirmed these findings. 6. The laboratory reports performing 317 urinalysis tests annually.</p>