

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 38D2039090	(X3) Date Survey Completed 12/16/2025
Name of Provider or Supplier Assure Drug Detection Llc	Street Address, City, State 607 Adams Ave, La Grande, OR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6092	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(iv)</p> <p>(e)(4)(iv) An approved corrective action plan is followed when any proficiency testing result is found to be unacceptable or unsatisfactory;</p> <p>This STANDARD is not met as evidenced by: Based on review of the 2024 American Proficiency Institute (API) proficiency test (PT) Chemistry - Miscellaneous - 2nd Event, and interview with the Technical Personnel (TP#2), it was revealed that the laboratory director (LD) failed to ensure corrective action documentation was performed for analytes with failed or unacceptable results for one (1) analyte. Findings include: 1. Upon review of the PT records presented for review during on site survey, it was revealed that the LD failed to ensure complete corrective action was documented for unacceptable PT results for Chemistry - Miscellaneous - 2nd Event in 2024. a. 2024 Chemistry- Miscellaneous - 2nd Event, no corrective action documentation for analytes: UDS Fentanyl (qual) - 67% 2. Request for the laboratory's procedure for unacceptable or failed PT results, none could be produced. 3. Interview with the TP #2 at 1:00pm on 12/16/2025 confirmed no corrective action was documented. 4. The laboratory reports performing 3000 Fentanyl tests annually.</p>